

## CE Declaration of Conformity

<b>Manufacturer/ Supplier Information:</b>	<b>Idaho Technology Inc.</b> 390 Wakara Way Salt Lake City, Utah 84108, USA Phone: 1-800-735-6544 <a href="mailto:it@idahotech.com">it@idahotech.com</a> <a href="http://www.idahotech.com">http://www.idahotech.com</a>
<b>Date of Conformity:</b>	27 JAN 06

**Idaho Technology declares that the products:**


<b>Product Name:</b>	R.A.P.I.D.® LT Instrument
<b>Description:</b>	Research instrument used for real-time PCR analysis of nucleic acids.

To which this declaration relates, meets the protection requirements laid down in Council Directive 89/336/EEC and 73/23/EEC on the alignment of the legal provisions of the Member States on electromagnetic compatibility (EMC).

To assess the product with regard to electromagnetic compatibility, the following relevant harmonized standards were applied:

<b>EN 502: 1998 (Class A)</b> Information technology equipment – Radio disturbance characteristics – Limits and methods of measurement
<b>EN 61010-1: 2001</b> Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirement
<b>EN 61010-2-010: 2003</b> Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2: Particular requirements for laboratory equipment for the heating of materials
<b>EN 55022: 1998</b> Information technology equipment – Radio disturbance characteristics – Limits and methods of measurement
<b>EN 55024: 1998</b> Information technology equipment – Immunity characteristics – Limits and methods of measurement
<b>EN 61000-3-2: 2000</b> Electromagnetic compatibility (EMC) – Limits – Limits for harmonic current emissions (Equipment input current ≤ 16 A per phase)
<b>EN 61000-3-3: 1995</b> Electromagnetic compatibility (EMC) – Limits – Section 3: Limitation of voltage fluctuations and flicker in low-voltage supply systems for equipment with rated current ≤ 16 A

We hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The unit complies with all essential requirements of the Directives.



**Marshall McCarty**  
Director of Quality Assurance and  
Regulatory Affairs



**Richard Andrew**  
Director of Instrument Production